


SUPPLEMENTARY INFORMATION

16(Py)-S-2-S-(Py)16-*d*₁₀

[M]²⁺: *m/z* 354.31 (C₄₄H₆₈D₁₀N₂S₂²⁺)



16-3-16-*d*₆₆

[M]²⁺: *m/z* 323.51 (C₃₉H₁₈D₆₆N₂²⁺)

m/z 388.61 (C₂₃H₁₈D₃₃N₂⁺)





Figure S1. Product ions monitored for the internal standards 16(Py)-S-2-S-(Py)16-*d*₁₀ and 16-3-16-*d*₆₆ during HILIC-LC-MS/MS analysis.

Figure S2:






Figure S2. Chromatograms of the gemini surfactants, 16(Py)-S-2-S-(Py)16, 16-3-16, 16(Py)-S-2-S-(Py)16-*d*₁₀, 16-3-16-*d*₆₆. [Panel **A** indicates the absence of any interference against the selective determination of all four compounds by showing the detection of no signals around the established HILIC-LC-MS/MS elution times of these compounds within "double blank" cell samples. **B**) chromatogram from an injection of only 16(Py)-S-2-S-(Py)16-*d*₁₀ and **C**) Chromatogram from an injection of only 16(Py)-S-2-S-(Py)16 at lower limit of quantitation, LLOQ. No cross-interference occurred between 16(Py)-S-2-S-(Py)16 and 16(Py)-S-2-S-(Py)16-*d*₁₀]. **D**) Chromatogram from an injection of only 16-3-16; and, **E**) Chromatogram from an injection of only 16-3-16-*d*₆₆. No cross-interference occurred between 16-3-16 and 16-3-16-*d*₆₆ since the injections produced signal for only the respective but not both compounds. In general, the chromatograms showed symmetry and smoothness for quantitative analysis as shown.

Figure S3:




Figure S3

Calibration curves for the 16-3-16 and 16(Py)-S-2-S-(Py)16 gemini surfactants. In both cases, the linear range was 50 – 5000 ng/[1 × 10⁶ cells], $r^2 \geq 0.99$.

...

Figure S4:






Figure S4. Chromatograms of the 16-3-16 gemini surfactant at LLOQ, ULOQ and LLOD. The relative response signal is shown for the analyte at: **A)** LLOQ and **B)** ULOQ in relation to the internal standard, 16-3-16-*d*₆₆, which was present at a constant concentration. **C)** Extracted chromatogram for 16-3-16 at LLOD.

...

Figure S5:




Figure S5. Contrasting analyte recovery efficiencies for Bligh/Dyer lipid extraction vs *octanol* extraction. Bligh/Dyer lipid extraction gave a better recovery (typically 98%) of the analyte and was the chosen liquid-liquid extraction method, departing from a recent report in which octanol extraction (70% efficiency) was used.

...

TABLES & TABLE CAPTIONS:

Table S1. Gemini surfactant mass concentrations equivalent to 3 mM molarity

Gemini surfactant	Mass concentration (\approx 3 mM)
12(Py)-S-2-S-(Py)12	2.24 mg/mL
14(Py)-S-2-S-(Py)14	2.40 mg/mL
16(Py)-S-2-S-(Py)16	2.57 mg/mL
18(Py)-S-2-S-(Py)18	2.74 mg/mL
16-3-16	2.22 mg/mL
16-7-16	2.39 mg/mL
12-3-12	1.88 mg/ml
12-4-12	1.92 mg/mL
12-8-12	2.09 mg/mL
12-10-12	2.18 mg/mL
12-12-12	2.26 mg/mL
12-16-12	2.43 mg/mL
18-3-18	2.39 mg/ml
18-7-18	2.55 mg/mL
18:1-3-18:1	2.37 mg/mL
18:1-6-18:1	2.50 mg/mL
12-7NH-12	2.05 mg/mL
Py-3-12	2.24 mg/mL

Table S2. Overview of the HILIC-LC-MS/MS method validation parameters

Analyte	16(Py)-S-2-S-(Py)16	16-3-16
Internal standard (IS)	16(Py)-S-2-S-(Py)16- <i>d</i> ₁₀	16-3-16- <i>d</i> ₆₆
Method description	LLE, HILIC-LC-MS/MS	LLE, HILIC-LC-MS/MS
QC range	150 – 4375 ng/mL	150 – 4375 ng/mL
Calibration curve range	50 – 5000 ng/mL	50 – 5000 ng/mL
Lower limit of detection (LLOD)	6 ng/mL	3 ng/mL
Average analyte recovery	97.7% (Bligh/Dyer method)	98.2% (Bligh/Dyer method)
Intra-day accuracy range	92 – 110%	92 – 110%
Intra-day precision range	3 – 11%	3 – 11%
Inter-day accuracy range	92 – 110%	92 – 110%
Inter-day precision range	3 – 11%	3 – 11%
Bench-top stability (hrs)	24	24
Autosampler stability (hrs)	48	48
Freeze-thaw stability (cycles)	3	3
Long-term stability (days)	105	105
Specificity (i.e., matrix effects)	No net matrix effects	No net matrix effects

...

Table S3. Recovery of 16-3-16 (analyte) from the aqueous cellular matrix

Theoretical concentration	Bligh/Dyer extraction			<i>Octanol</i> extraction		
	Extracted cells (mean ± SD, ng/mL)	Spiked extracted cells (mean ± SD, ng/mL)	Recovery (%)	Extracted cells (mean ± SD, ng/mL)	Spiked extracted cells (mean ± SD, ng/mL)	Recovery (%)
16-3-16						
50 ng/mL (LLOQ)	48.24 ± 5.73	49.38 ± 3.74	97.7	36.38 ± 5.02	52.21 ± 4.74	69.7
150 ng/mL (LQC)	147.19 ± 13.61	150.53 ± 15.31	97.8	114.51 ± 8.72	159.26 ± 11.38	71.9
375 ng/mL (MQC)	371.57 ± 51.33	383.46 ± 47.32	96.9	277.87 ± 33.16	385.06 ± 37.92	72.2
4375 ng/mL (HQC)	4359.45 ± 359.41	4368.17 ± 212.57	100.2	3210.64 ± 179.11	4528.22 ± 182.33	70.9

...

Table S4. Intra- and inter-day accuracy and precision in the analysis of 16-3-16

Quality Control		Observed concentration (mean \pm SD, ng/mL)	Accuracy (%)	Precision (%RSD)
INTRA-DAY:				
	<u>Analysis Day (#)</u>			
LLOQ: 50 ng/mL	Day 1	47.45 \pm 4.22	94.9	8.9
	Day 2	49.61 \pm 3.77	99.2	7.6
	Day 3	50.19 \pm 3.31	100.4	6.6
LQC: 150 ng/mL	Day 1	160.95 \pm 14.65	107.3	9.1
	Day 2	159.45 \pm 13.39	106.3	8.4
	Day 3	163.78 \pm 14.58	109.2	8.9
MQC: 375 ng/mL	Day 1	343.88 \pm 34.73	91.7	10.1
	Day 2	392.62 \pm 33.37	104.7	8.5
	Day 3	336.75 \pm 32.33	89.8	9.6
HQC: 4375 ng/mL	Day 1	4480.01 \pm 456.96	102.4	10.2
	Day 2	4536.88 \pm 426.47	103.7	9.4
	Day 3	4738.13 \pm 355.36	108.3	7.5
INTER-DAY:				
	<u>Concentration</u>			
LLQC	50 ng/mL	55.05 \pm 5.84	110.1	10.6
LQC	150 ng/mL	145.955 \pm 11.09	97.3	7.6
MQC	375 ng/mL	393.03 \pm 34.19	104.8	8.7
HQC	4375 ng/mL	4475.63 \pm 434.14	102.3	9.7

...

Table S5. Stability of the 16-3-16 gemini surfactant within the sample matrix

Quality Control	Storage condition/ period	Observed concentrations (mean \pm SD, ng/mL)	Accuracy (%)	Precision (%RSD)
LQC: 150 ng/mL	0 h	150.31 \pm 11.57	100.2	7.7
	24 h on bench top	164.42 \pm 10.69	109.6	6.5
	Three freeze/thaw cycles	135.45 \pm 14.22	90.3	10.5
	-20 °C for 105 days (LT)	134.85 \pm 8.23	89.9	6.1
	48 h extract in autosampler	144.15 \pm 13.84	96.1	9.6
MQC: 375 ng/mL	0 h	362.63 \pm 34.45	96.7	9.5
	24 h on bench top	353.62 \pm 34.31	94.3	9.7
	Three freeze/thaw cycles	338.63 \pm 32.17	90.3	9.5
	-20 °C for 105 days (LT)	345.38 \pm 36.26	92.1	10.5
	48 h extract in autosampler	399.75 \pm 37.58	106.6	9.4
HQC: 4375 ng/mL	0 h	4147.51 \pm 443.78	94.8	10.7
	24 h on bench top	4379.38 \pm 381.01	100.1	8.7
	Three freeze/thaw cycles	4038.13 \pm 415.93	92.3	10.3
	-20 °C for 105 days (LT)	4239.38 \pm 390.02	96.9	9.2
	48 h extract in autosampler	4392.51 \pm 347.01	100.4	7.9

LT (long term): -20 °C for 105 days

...